

	Technical Data												
Use in	- Pharmaceutical Industry - Cosmetic Industry - Food Industry - For industrial, laboratory & research applications only												
Use for	- Test for specified micro-organisms - Preparation of test strains - Diluent for sample preparation - Identification and growth of aerobic micro-organisms												
Typical composition per liter	Basic medium according to Ph. Eur. 2.6.1 and USP <71> <table><tr><td>Casein peptone</td><td>17 g</td><td>Glucose x H₂O</td><td>2.5 g</td></tr><tr><td>Soy peptone</td><td>3 g</td><td>K₂HPO₄</td><td>2.5 g</td></tr><tr><td>NaCl</td><td>5 g</td><td></td><td></td></tr></table> This medium can be adjusted / or supplemented according to the performance criteria required.	Casein peptone	17 g	Glucose x H ₂ O	2.5 g	Soy peptone	3 g	K ₂ HPO ₄	2.5 g	NaCl	5 g		
Casein peptone	17 g	Glucose x H ₂ O	2.5 g										
Soy peptone	3 g	K ₂ HPO ₄	2.5 g										
NaCl	5 g												
Filling volume	100 mL												
Bottle format	220 mL screw cap - Type II glass - Bottle opening about 31 mm - Colour of cap: green - GL40 screw cap with 2 integrated septa												
Bottles per tray	12 bottles on a plastic tray wrapped with shrink foil												
Shelf life	12 months from production date												
Storage conditions	- Recommended storage temperature: 2 – 25 °C - Should be stored at temperatures as stable as possible - Avoid prolonged exposure to direct sunlight												
Label	- On the side - Contain autoclave indicator												
Label information	- Product name: TSB 100 mL - Expiry date: YYYYMMDD ➔ MMM in letters (e.g.: 2026Nov04) - Lot-number - Individual number - Barcode												
Barcode	2-dimensional (data matrix), 20 digits: - Digits 1 - 3: Art.-No. - Digits 4 - 9: Lot-Number - Digits 10 - 14: Individual-Number - Digits 15 - 20: Date (YYMMDD)												

	Technical Data
Delivery	• Temperature controlled delivery on request • For shipments of larger amounts plastic pallets in Euro-size can be used
Bottle information	• Label contains autoclaving indicator (brown → green) • Bottles are incubated at 25 – 35 °C for at least 48 hours after autoclaving and then packed • Bottles are not touched anymore by hand after autoclaving
Place of production	PharmaMedia Dr. Müller GmbH Gustav-Throm-Str. 1, 69181 Leimen - Germany

Quality control, Certificates					
Certificates	Each batch of product can be obtained with a certificate of analysis (CoA):				
	Physico-chemical test parameters:				
	Appearance	Clear, yellowish			
	pH value	7.1 – 7.5			
	Filling volume	98 – 104 mL			
	Growth Promotion test: 10 – 100 CFU*				
	<i>S. aureus</i>	ATCC 6538	30 – 35 °C	1 day	Good growth
	<i>E. coli</i>	ATCC 8739	30 – 35 °C	1 day	Good growth
	<i>P. paraeruginosa</i>	ATCC 9027	30 – 35 °C	1 day	Good growth
	<i>B. spizizenii</i>	ATCC 6633	30 – 35 °C	1 day	Good growth
	<i>C. albicans</i>	ATCC 10231	20 – 25 °C	3-5 days	Good growth
	<i>A. brasiliensis</i>	ATCC 16404	20 – 25 °C	3-5 days	Good growth
Sterility control					
≥ 7 days at 30 – 35 °C, no growth					
*In case of direct inoculation and incubation of the bottle, please check the recommendations for use on next page.					
Certificate of origin	All media lots produced by PMM can be obtained with a Certificate of Origin (CoO). All animal derived raw materials are specified as follows: • Raw material • Tissue • Animal source • Country of origin • Infectivity category (acc. to TSE guideline: EMA/410/01 current version)				

Quality control, Certificates	
BSE policy	In compliance with the current note for guidance on minimizing the risk of transmitting animal spongiform encephalopathy via human or veterinary medicinal products, we check the CoO of raw material in respect to the specified animal source, the country of origin and the infectivity category. We neither store or process ruminant raw materials obtained from high infectivity tissues (IA) nor ruminant raw materials whose animal source originates from countries or regions with an undetermined risk (cat C/GBR IV).
Temperature stress	Art. 500.B100 has been exposed to temperature stress conditions (3 days at 30-35 °C) and has passed shelf-life testing at least 30 days after the assigned expiry date. Shelf-life testing comprises all regular tests which are part of the normal release test of this article except for sterility control (see CoA).

Recommendations for use	
Release of negative pressure in media bottles	During the autoclaving process, chemical reactions inside the bottles may result in a slight vacuum. Please ensure that the vacuum is released without contaminating the bottle. Ideally, the vacuum is released by puncturing the septum with an aeration needle equipped with a sterile filter prior to opening a bottle.
Aeration for direct inoculation	During the autoclaving process, most of the oxygen inside the bottles is consumed in a chemical reaction, thus resulting in a medium not suitable for direct inoculation of aerobic microorganisms. When using such bottle for the growth of aerobic microorganisms without opening, please ensure that the bottle is aerated throughout the complete incubation process by an aeration needle equipped with a sterile filter. Additionally, air filtered through a sterile filter may be pressed into the bottle using a syringe. Example for aeration: For bottles containing larger volumes of medium, e.g. TSB in 500 or 1000 mL, puncture the bottle lid (stopper) by a cannula of at least 1.6 mm diameter equipped with a sterile filter. Equilibrate with the cannula for not less than three days at 20 to 25 °C prior to inoculation.

Safety Data	
Toxic ingredients	• None
Basic composition	• See typical composition
Solvent content	• None
Safety data sheet required	• Not mandatorily required