Technical Data Sheet – Item No. 588.A100

MacConkey Bouillon acc. Ph. Eur./USP

Version 03
Effective date: 18-11-2026

	Technical Data
Use in	• Pharmaceutical Industry • For industrial, laboratory & research applications only
Use for	• Microbial examination of non-sterile products • Test for specified micro-organisms • Test for *E. coli*
Typical composition per liter	Basic medium according to Ph. Eur. 2.6.13 und USP <62> Pancreatic digest of gelatine 20 g — Dehydrated ox bile 5 g Lactose monohydrate 10 g — Bromocresol purple 10 mg This medium can be adjusted and/or supplemented according to the performance criteria required.
Filling volume	• 100 mL
Bottle format	• 120 mL screw cap • Type III glass • Bottle opening about 20 mm • Colour of cap: blue • PP28 screw cap with 2 integrated septa
Bottles per tray	• 20 bottles on a plastic tray wrapped with shrink foil
Shelf life	• 9 months from production date
Storage conditions	• Recommended storage temperature: 2-25 °C • Store protected from light exposure
Label	• On the side, at the bottom • Contain autoclave indicator
Label information	• Product name: MacC.-B. 100 mL • Expiry date: YYYYMMMDD ➔ MMM in letters (e.g.: 2027Nov04) • Lot-number • Individual number • Barcode
Barcode	• 2-dimensional (data matrix), 20 digits: • Digits 1-3: Art.-No. • Digits 4-9: Lot-Number • Digits 10-14: Individual-Number • Digits 15-20: Date (YYMMDD)

Pharmamedia Dr. Müller GmbH | Gustav-Throm-Str.1 | 69181 Leimen | Germany
Phone: +49 (0) 6224 928070 | Email address: info@pmm-leimen.de | www.pmm-leimen.de

Technical Data Sheet – Item No. 588.A100

MacConkey Bouillon acc. Ph. Eur./USP

Version 03
Effective date: 18-11-2026

	Technical Data
Delivery	• Temperature controlled delivery on request • For shipments of larger amounts plastic pallets in Euro-size can be used
Bottle information	• Label contains autoclaving indicator (brown → green) • Bottles are incubated at 25 - 35 °C for at least 48 hours after autoclaving and then packed • Bottles are not touched any more by hand after autoclaving
Place of production	PharmaMedia Dr. Müller GmbH Gustav-Throm-Str. 1, 69181 Leimen - Germany

Quality control, Certificates

Certificates

Every batch of product can be obtained with a certificate of analysis (CoA):

Physico-chemical test parameters:				
Appearance	Clear, purple			
pH value	7.1 - 7.5			
Filling volume	98 - 104 mL			
Growth Promotion test: 10-100 CFU*				
E. coli	ATCC 8739	42-44 °C	20-24 h	Good growth
Test for inhibition: 100-1,000 CFU				
S. aureus	ATCC 6538	42-44 °C	48-52 h	Inhibition
Sterility control				
≥ 7 days at 30-35 °C, no growth				

*In case of a direct inoculation and incubation in the bottle, please check the recommendations for use on next page.

Certificate of origin

All media lots produced by PMM can be obtained with a Certificate of Origin (CoO). All animal derived raw materials are specified as follows:

- Raw material
- Tissue
- Animal source
- Country of origin
- Infectivity category (acc. to TSE guideline: EMA/410/01 current version)

Non-animal-derived raw materials are not listed in the CoO; raw materials listed on the CoA in the typical composition section but not listed on the CoO are of non-animal origin.

Pharmamedia Dr. Müller GmbH | Gustav-Throm-Str.1 | 69181 Leimen | Germany
Phone: +49 (0) 6224 928070 | Email address: info@pmm-leimen.de | www.pmm-leimen.de

	Quality control, Certificates
BSE policy	In compliance with the current note for guidance on minimizing the risk of transmitting animal spongiform encephalopathy via human or veterinary medicinal products, we check the CoO of raw material in respect to the specified animal source, the country of origin and the infectivity category. We neither store or process ruminant raw materials obtained from high infectivity tissues (IA) nor ruminant raw materials whose animal source originates from countries or regions with an undetermined risk (cat C/GBR IV).

	Recommendations for use
Release of negative pressure in media bottles	During the autoclaving process, chemical reactions inside the bottles may result in a slight vacuum. Please ensure that the vacuum is released without contaminating the bottle. Ideally, the vacuum is released by puncturing the septum with an aeration needle equipped with a sterile filter prior to opening a bottle.
Aeration for direct inoculation	During the autoclaving process, most of the oxygen inside the bottles is consumed in a chemical reaction, thus resulting in a medium not suitable for direct inoculation of aerobic microorganisms. When using such bottle for the growth of aerobic microorganisms without opening, please ensure that the bottle is aerated throughout the complete incubation process by an aeration needle equipped with a sterile filter. Additionally, air filtered through a sterile filter may be pressed into the bottle using a syringe.

	Safety Data
Toxic ingredients	• None
Basic composition	• See typical composition
Solvent content	• None
Safety data sheet required	• Not mandatorily required

Pharmamedia Dr. Müller GmbH | Gustav-Throm-Str.1 | 69181 Leimen | Germany
Phone: +49 (0) 6224 928070 | Email address: info@pmm-leimen.de | www.pmm-leimen.de