	Technical Data Sheet				
Use in	<ul> <li>Pharmaceutical Industry</li> <li>For industrial, laboratory &amp; research applications only</li> <li>Basic medium according to EP 2.6.13 and USP &lt;62&gt;</li> </ul>				
Use for	<ul> <li>Examination of non-sterile products</li> <li>Test for specified micro-organisms</li> <li>Test for Bile Tolerant Gram-negative bacteria</li> <li>Application:         Typically, 10 g of product are diluted in 90 mL Tryptic Soy Broth (TSB, e.g., artNo. 501.B090), mixed and incubated at 20-25 °C for 2-5 h. The amount corresponding to 1 g of product is transferred to Mossel Broth and incubated at 30-35 °C for 24-48 h. An aliquot is subcultured on VRBD Agar and incubated at 30-35 °C for 18-24 h. Growth of colonies indicate the presence of Bile Tolerant Gram-Negative Bacteria.     </li> </ul>				
Typical composition per liter	Yeast extract 3 g Bile salts 1.5 g Pancreatic digest of casein 7 g Neutral red 30 mg Glucose monohydrate 10 g Crystal Violet 2 mg NaCl 5 g Agar 15 g  This medium can be adjusted / or supplemented according to the performance criteria required.				
Irradiation	Not irradiated				
Filling volume	• 23-26 mL				
Packaging	<ul> <li>Single bagged, staples of 10 plates</li> <li>Transparent</li> <li>High barrier foil against desiccation</li> <li>6 staples of 10 plates per packaging unit</li> <li>Temperature isolated handle-bag in the cardboard-boxes</li> </ul>				
Plates per box	60 plates (6 staples with 10 plates each)				
Shelf life	8 months from production date				
Storage conditions	<ul> <li>Recommended storage temperature: 15-25 °C</li> <li>Should be stored at temperatures as stable as possible</li> <li>Store protected from light exposure</li> <li>Before use: it is recommended to keep the plates upright (agar on the lower part, lid on the upper part) to avoid formation of extra condensation</li> <li>After use: it is recommended to keep the plates upside down (agar on the upper part, lid on the lower part) to reduce the risk of accumulation of condensation during incubation which can affect colony formation</li> </ul>				
Label	On the side of the bottom part of the dish				



	Technical Data Sheet			
Label information	<ul> <li>Product name: VRBD</li> <li>Expiry date: YYYYMMMDD → MMM in letters (e.g.: 2023Nov04)</li> <li>Lot-number</li> <li>Individual number</li> <li>Barcode</li> </ul>			
Barcode	<ul> <li>2-dimensional (data matrix), 20 digits:</li> <li>Digits 1-3: ArtNo.</li> <li>Digits 4-9: Lot-Number</li> <li>Digits 10-14: Individual-Number</li> <li>Digits 15-20: Date (YYMMDD)</li> </ul>			
Delivery	<ul> <li>Temperature controlled delivery on request</li> <li>For shipments of larger amounts plastic pallets in Euro-size can be used</li> </ul>			
Petri dish	<ul> <li>Locking lid 90 mm plate, made from Polystyrene</li> <li>Long incubations possible – due to high filling volume</li> <li>Long expositions possible – due to specific design of plate</li> <li>Incubations in vent and closed position possible</li> </ul>			
Lid positions	<ul> <li>All plates are delivered in the non-locked position</li> <li>The plate contains 2 locked positions. If turning the lid clockwise the locked positions are in the following order:</li> <li>1. Vent position</li> <li>2. Closed position</li> </ul>			
Aerobic incubation	<ul> <li>The closed position provides ideal incubation conditions for aerobic microorganisms and limits the dehydration of the agar during incubation</li> <li>For long incubation of aerobic microorganisms, the closed position is recommended</li> <li>To lock the lid in the closed position turn the lid clockwise into the final stop position</li> </ul>			
Place of production	PharmaMedia Dr. Müller GmbH Gustav-Throm-Str. 1, 69181 Leimen - Germany			



	Quality control, Certificates						
	Each lot of product can be obtained with a certificate of analysis (CoA):						
	Appearance	Physico-chemical test parameters:					
	pH value	Clear, slightly reddish 7.2 – 7.6					
	Filling volume	23 – 26 mL					
	Growth Promoti			40.40 %	50.2000/		
	E.coli	ATCC 8739 ATCC 9027	30-35 °C 30-35 °C	16-18 h 16-18 h	50-200% 50-200%		
	P.paraeruginosa	ATCC 9021	30-35 C	10-1011	30-200%		
	Inhibition test: 1	00-1,000 CFU					
Certificates	S.aureus	ATCC 6538	30-35 °C	24-28 h	No growth		
Certificates							
	Indicative prope			40.04 5			
	E.coli Good growth, red	ATCC 8739	30-35 °C	18-24 h			
	P.paraeruginosa		30-35 °C				
	Good growth, col	•			on zone		
	Cood growin, con	oned to print ee	iornoo wiinou	t proorpitati	011 20110		
	Sterility control				Conform		
	* ^	D 0 0 40 H0D 40	0>	4-4: 44 /			
	recovery rate ag				determination of		
	, ,			,			
		qualitative comparison with a previous approved batch is requested.  However, during quality control at PMM recovery rate data are determined					
	in reference to a previously released batch of the product.						
	In compliance	with the current	note for guida	ance on mir	nimizina the risk of		
	<ul> <li>In compliance with the current note for guidance on minimizing the risk of transmitting animal spongiform encephalopathy via human or veterinary</li> </ul>						
	medicinal products, we check the CoO of raw material in respect to the						
BSE policy	specified animal source, the country of origin and the infectivity category.						
DOL policy	We neither store or process ruminant raw materials obtained from high						
	infectivity tissues (IA) nor ruminant raw materials whose animal source						
	originates from countries or regions with an undetermined risk (cat C/GBR IV).						
	,						
					conditions (3 days		
Temperature stress	at 2-8 °C as well as 3 days at 30-35 °C) and has passed shelf-life testing						
	at least 210 days after production date. Shelf-life testing comprise all regular tests which are part of the normal release test of this article except						
	for sterility control (see CoA).						
		(=== 00.1).					

. <u>.</u>	Safety Data
Toxic ingredients	• None



	Safety Data	
Basic composition	See typical composition	
Solvent content	• None	
Safety data sheet required	Not mandatorily required	