

	Technical Data Sheet			
Use in	<ul style="list-style-type: none">Pharmaceutical IndustryFor industrial, laboratory & research applications onlyBasic medium according to EP 2.6.13 and USP <62>			
Use for	<ul style="list-style-type: none">Examination of non-sterile productsTest for specified micro-organismsTest for <i>Salmonella</i> spp. <p>Application: Typically, 10 g of product are diluted in 90 mL Tryptic Soy Broth (TSB, e.g., art.-No. 501.B090), mixed and incubated at 30-35 °C for 18-24 h. 0.1 mL is transferred to Rappaport-Vassiliadis Broth and incubated at 30-35 °C for 18-24 h. An aliquot is subcultured on XLD Agar and incubated at 30-35 °C for 18-48 h. Growth of reddish or red colonies, with or without black centre indicates the presence of <i>Salmonella</i>. In case such colonies are detected an identification test needs to prove the presence/absence of <i>Salmonella</i>.</p>			
Typical composition per liter	Xylose	3.5 g	Yeast extract	3 g
	L-Lysine	5 g	NaCl	5 g
	Na-deoxycholate	2.5 g	Ferric ammonium citrate	0.8 g
	Lactose x H ₂ O	7.5 g	Phenol red	80 mg
	Sucrose	7.5 g	Agar	13.5 g
	Na-thiosulfate	6.8 g		
	This medium can be adjusted / or supplemented according to the performance criteria required.			
Irradiation	<ul style="list-style-type: none">Not irradiated			
Filling volume	<ul style="list-style-type: none">23-26 mL			
Packaging	<ul style="list-style-type: none">Single bagged, staples of 10 platesTransparentHigh barrier foil against desiccation6 staples of 10 plates per packaging unitTemperature isolated handle-bag in the cardboard-boxes			
Plates per box	<ul style="list-style-type: none">60 plates (6 staples with 10 plates each)			
Shelf life	<ul style="list-style-type: none">4 months from production date			

	Technical Data Sheet
Storage conditions	<ul style="list-style-type: none"> Recommended storage temperature: 15-25 °C Should be stored at temperatures as stable as possible Store protected from light exposure Before use: it is recommended to keep the plates upright (agar on the lower part, lid on the upper part) to avoid formation of extra condensation After use: it is recommended to keep the plates upside down (agar on the upper part, lid on the lower part) to reduce the risk of accumulation of condensation during incubation which can affect colony formation
Label	<ul style="list-style-type: none"> On the side, at the bottom
Label information	<ul style="list-style-type: none"> Product name: XLD Expiry date: YYYYMMDD → MMM in letters (e.g.: 2023Nov04) Lot-number Individual number Barcode
Barcode	<ul style="list-style-type: none"> 2-dimensional (data matrix), 20 digits: Digits 1-3: Art.-No. Digits 4-9: Lot-Number Digits 10-14: Individual-Number Digits 15-20: Date (YYMMDD)
Delivery	<ul style="list-style-type: none"> Temperature controlled delivery on request For shipments of larger amounts plastic pallets in Euro-size can be used
Petri dish	<ul style="list-style-type: none"> Locking lid 90 mm plate, made from polystyrene Long incubations possible – due to high filling volume Long expositions possible – due to specific design of plate Incubations in vent and closed position possible
Lid positions	<ul style="list-style-type: none"> All plates are delivered in the non-locked position The plate contains 2 locked positions. If turning the lid clockwise the locked positions are in the following order: <ol style="list-style-type: none"> Vent position Closed position
Aerobic incubation	<ul style="list-style-type: none"> The closed position provides ideal incubation conditions for aerobic microorganisms and limits the dehydration of the agar during incubation For long incubation of aerobic microorganisms, the closed position is recommended To lock the lid in the closed position turn the lid clockwise into the final stop position
Production	<p>Production of selective media is made at the beginning of every quarter. To get the longest shelf life possible, we do recommend to place orders <u>for delivery</u> at this period of time.</p>

Technical Data Sheet	
----------------------	--

Place of production	PharmaMedia Dr. Müller GmbH
	Gustav-Throm-Str. 1, 69181 Leimen - Germany

Quality control, Certificates	
-------------------------------	--

Certificates	Each lot of product can be obtained with a certificate of analysis (CoA):
--------------	---

Certificates	Physico-chemical test parameters:				
--------------	--	--	--	--	--

Certificates	Appearance	Clear, red, possibly with white crystals			
--------------	------------	--	--	--	--

Certificates	pH value	7.2 – 7.6			
--------------	----------	-----------	--	--	--

Certificates	Filling volume	23 – 26 mL			
--------------	----------------	------------	--	--	--

Certificates	Growth Promotion test: 10-100 CFU*				
--------------	---	--	--	--	--

Certificates	S.Typhimurium	ATCC 14028	30-35 °C	16-18 h	50-200%
--------------	---------------	------------	----------	---------	---------

Certificates	S.Abony	NCTC 6017	30-35 °C	16-18 h	50-200%
--------------	---------	-----------	----------	---------	---------

Certificates	Inhibition test: 10,000-100,000 CFU				
--------------	--	--	--	--	--

Certificates	S.aureus	ATCC 6538	30-35 °C	48-52 h	No growth
--------------	----------	-----------	----------	---------	-----------

Certificates	Indicative properties: 10-100 CFU				
--------------	--	--	--	--	--

Certificates	S.Typhimurium	ATCC 14028	30-35 °C	18-24 h	
--------------	---------------	------------	----------	---------	--

Certificates	Colorless to pale pink colonies, colonies mostly with black center, medium red				
--------------	--	--	--	--	--

Certificates	S.Abony	NCTC 6017	30-35 °C	18-24 h	
--------------	---------	-----------	----------	---------	--

Certificates	Colorless to pale pink colonies, colonies mostly with black center, medium red				
--------------	--	--	--	--	--

Certificates	Sterility control				Conform
--------------	--------------------------	--	--	--	---------

Certificates	<p>*According to EP 2.6.13 USP<62>, no quantitative test (determination of recovery rate against a non-selective reference) is required – only the qualitative comparison with a previous approved batch is requested. However, during quality control at PMM recovery rate data are determined in reference to a previously released batch of the product.</p>				
--------------	---	--	--	--	--

BSE policy	<ul style="list-style-type: none"> In compliance with the current note for guidance on minimizing the risk of transmitting animal spongiform encephalopathy via human or veterinary medicinal products, we check the CoO of raw material in respect to the specified animal source, the country of origin and the infectivity category. We neither store or process ruminant raw materials obtained from high infectivity tissues (IA) nor ruminant raw materials whose animal source originates from countries or regions with an undetermined risk (cat C/GBR IV).
------------	---

Quality control, Certificates	
Temperature stress	<ul style="list-style-type: none"> Art. 470.0060 has been exposed to temperature stress conditions (3 days at 2-8 °C as well as 3 days at 30-35 °C) and has passed shelf-life testing at least 30 days after the assigned expiry date. Shelf-life testing comprise all regular tests which are part of the normal release test of this article except for sterility control (see CoA).

Safety Data	
Toxic ingredients	<ul style="list-style-type: none"> None
Basic composition	<ul style="list-style-type: none"> See typical composition
Solvent content	<ul style="list-style-type: none"> None
Safety data sheet required	<ul style="list-style-type: none"> Not mandatorily required