

<b>Technical Data Sheet</b>																	
<b>Use in</b>	<ul style="list-style-type: none"> <li>Pharmaceutical industry</li> <li>For industrial, laboratory &amp; research applications only</li> </ul>																
<b>Use for</b>	<ul style="list-style-type: none"> <li>Sterility test</li> <li>Identification and growth of fastidious anaerobic micro-organisms as well as aerobic micro-organisms</li> </ul>																
<b>Typical composition per liter</b>	<p>Basic medium according to Ph. Eur. 2.6.1 and USP &lt;71&gt;</p> <table style="width: 100%; border-collapse: collapse;"> <tbody> <tr> <td style="width: 33%;">Casein peptone</td> <td style="width: 16.5%;">15 g</td> <td style="width: 33%;">Glucose-D(+) x H<sub>2</sub>O</td> <td style="width: 16.5%;">5.5 g</td> </tr> <tr> <td>Yeast extract</td> <td>5 g</td> <td>Na-Thioglycolate</td> <td>0.5 g</td> </tr> <tr> <td>NaCl</td> <td>2.5 g</td> <td>Resazurine</td> <td>1 mg</td> </tr> <tr> <td>Agar/ Gel Agent</td> <td>0.2 g</td> <td>L-Cysteine HCl</td> <td>0.5 g</td> </tr> </tbody> </table> <p>This medium can be adjusted / or supplemented according to the performance criteria required.</p>	Casein peptone	15 g	Glucose-D(+) x H <sub>2</sub> O	5.5 g	Yeast extract	5 g	Na-Thioglycolate	0.5 g	NaCl	2.5 g	Resazurine	1 mg	Agar/ Gel Agent	0.2 g	L-Cysteine HCl	0.5 g
Casein peptone	15 g	Glucose-D(+) x H <sub>2</sub> O	5.5 g														
Yeast extract	5 g	Na-Thioglycolate	0.5 g														
NaCl	2.5 g	Resazurine	1 mg														
Agar/ Gel Agent	0.2 g	L-Cysteine HCl	0.5 g														
<b>Filling volume</b>	<ul style="list-style-type: none"> <li>200 mL</li> </ul>																
<b>Bottle format</b>	<ul style="list-style-type: none"> <li>220 mL screw cap</li> <li>Type II glass</li> <li>Bottle opening about 31 mm</li> <li>Colour of cap: blue</li> <li>GL40 screw cap with 2 integrated septa</li> </ul>																
<b>Bottles per tray</b>	<ul style="list-style-type: none"> <li>12 bottles on a plastic tray wrapped with shrink foil</li> </ul>																
<b>Shelf life</b>	<ul style="list-style-type: none"> <li>8 months from production date</li> </ul>																
<b>Storage conditions</b>	<ul style="list-style-type: none"> <li>Recommended storage temperature: 2 - 25 °C</li> <li>Should be stored at temperatures as stable as possible</li> <li>Store protected from light exposure</li> </ul>																
<b>Label</b>	<ul style="list-style-type: none"> <li>On the side</li> <li>Contain autoclave indicator</li> </ul>																
<b>Label information</b>	<ul style="list-style-type: none"> <li>Product name: FTM clear 200 mL</li> <li>Expiry date: YYYYMMDD → MMM in letters (e.g.: 2026Nov04)</li> <li>Lot-number</li> <li>Individual number</li> <li>Barcode</li> </ul>																
<b>Barcode</b>	<ul style="list-style-type: none"> <li>2-dimensional (data matrix), 20 digits:</li> <li>Digits 1-3: Art.-No.</li> <li>Digits 4-9: Lot-Number</li> <li>Digits 10-14: Individual-Number</li> <li>Digits 15-20: Date (YYMMDD)</li> </ul>																
<b>Delivery</b>	<ul style="list-style-type: none"> <li>Temperature controlled delivery on request</li> <li>For shipments of larger amounts plastic pallets in Euro-size can be used</li> </ul>																

<b>Technical Data Sheet</b>	
<b>Bottle information</b>	<ul style="list-style-type: none"> <li>• Label contains autoclaving indicator (brown → green)</li> <li>• Bottles are incubated at 25 – 35 °C for at least 48 hours after autoclaving and then packed</li> <li>• Bottles are not touched any more by hand after autoclaving</li> </ul>
<b>Place of production</b>	PharmaMedia Dr. Müller GmbH Gustav-Throm-Str. 1, 69181 Leimen - Germany

<b>Quality control, Certificates</b>																																																																							
<b>Certificates</b>	<p>Every batch of product can be obtained with a certificate of analysis (CoA):</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="5" style="text-align: left;"><b>Physico-chemical test parameters:</b></th> </tr> </thead> <tbody> <tr> <td style="width: 25%;">Appearance</td> <td colspan="4">Clear, yellowish</td> </tr> <tr> <td>pH value</td> <td colspan="4">6.9 – 7.3</td> </tr> <tr> <td>Filling volume</td> <td colspan="4">196 – 208 mL</td> </tr> <tr> <td colspan="5"> </td> </tr> <tr> <th colspan="5" style="text-align: left;"><b>Growth Promotion test: 10 – 100 CFU</b></th> </tr> <tr> <td><i>S. aureus</i></td> <td>ATCC 6538</td> <td>30 – 35 °C</td> <td>≤ 3 days</td> <td>Good growth</td> </tr> <tr> <td><i>P. paraeruginosa</i></td> <td>ATCC 9027</td> <td>30 – 35 °C</td> <td>≤ 3 days</td> <td>Good growth</td> </tr> <tr> <td><i>C. sporogenes</i></td> <td>ATCC 11437</td> <td>30 – 35 °C</td> <td>≤ 3 days</td> <td>Good growth</td> </tr> <tr> <td><i>C. sporogenes</i></td> <td>ATCC 19404</td> <td>30 – 35 °C</td> <td>≤ 3 days</td> <td>Good growth</td> </tr> <tr> <td><i>C. acnes</i></td> <td>ATCC 11827</td> <td>30 – 35 °C</td> <td>≤ 5 days</td> <td>Good growth</td> </tr> <tr> <td colspan="5"> </td> </tr> <tr> <th colspan="5" style="text-align: left;"><b>Sterility control</b></th> </tr> <tr> <td colspan="5">≥ 7 days at 30 – 35 °C, no growth</td> </tr> </tbody> </table>	<b>Physico-chemical test parameters:</b>					Appearance	Clear, yellowish				pH value	6.9 – 7.3				Filling volume	196 – 208 mL									<b>Growth Promotion test: 10 – 100 CFU</b>					<i>S. aureus</i>	ATCC 6538	30 – 35 °C	≤ 3 days	Good growth	<i>P. paraeruginosa</i>	ATCC 9027	30 – 35 °C	≤ 3 days	Good growth	<i>C. sporogenes</i>	ATCC 11437	30 – 35 °C	≤ 3 days	Good growth	<i>C. sporogenes</i>	ATCC 19404	30 – 35 °C	≤ 3 days	Good growth	<i>C. acnes</i>	ATCC 11827	30 – 35 °C	≤ 5 days	Good growth						<b>Sterility control</b>					≥ 7 days at 30 – 35 °C, no growth				
<b>Physico-chemical test parameters:</b>																																																																							
Appearance	Clear, yellowish																																																																						
pH value	6.9 – 7.3																																																																						
Filling volume	196 – 208 mL																																																																						
<b>Growth Promotion test: 10 – 100 CFU</b>																																																																							
<i>S. aureus</i>	ATCC 6538	30 – 35 °C	≤ 3 days	Good growth																																																																			
<i>P. paraeruginosa</i>	ATCC 9027	30 – 35 °C	≤ 3 days	Good growth																																																																			
<i>C. sporogenes</i>	ATCC 11437	30 – 35 °C	≤ 3 days	Good growth																																																																			
<i>C. sporogenes</i>	ATCC 19404	30 – 35 °C	≤ 3 days	Good growth																																																																			
<i>C. acnes</i>	ATCC 11827	30 – 35 °C	≤ 5 days	Good growth																																																																			
<b>Sterility control</b>																																																																							
≥ 7 days at 30 – 35 °C, no growth																																																																							
<b>Certificate of origin</b>	<p>All media lots produced by PMM can be obtained with a Certificate of Origin (CoO). All animal derived raw materials are specified as follows:</p> <ul style="list-style-type: none"> <li>• Raw material</li> <li>• Tissue</li> <li>• Animal source</li> <li>• Country of origin</li> <li>• Infectivity category (acc. to TSE guideline: EMA/410/01 current version)</li> </ul>																																																																						
<b>BSE policy</b>	<p>In compliance with the current note for guidance on minimizing the risk of transmitting animal spongiform encephalopathy via human or veterinary medicinal products, we check the CoO of raw material in respect to the specified animal source, the country of origin and the infectivity category. We neither store or process ruminant raw materials obtained from high infectivity tissues (IA) nor ruminant raw materials whose animal source originates from countries or regions with an undetermined risk (cat C/GBR IV).</p>																																																																						

	Recommendations for use
<b>Release of negative pressure in media bottles</b>	During the autoclaving process, chemical reactions inside the bottles may result in a slight vacuum. Please ensure that the vacuum is released without contaminating the bottle. Ideally, the vacuum is released by puncturing the septum with an aeration needle equipped with a sterile filter prior to opening a bottle.
<b>Resazurine</b>	Resazurine is an indicator for anaerobic conditions. The indicator is colourless under anaerobic condition and turns into pink colour under aerobic conditions. Due to the composition of the medium 2 relatively stable zones can be clearly separated: 1. an aerobic, pink zone on the top 2. an anaerobic, yellowish zone on the bottom. Acc. to Ph. Eur./USP not more than the upper third of the medium should be pink coloured. If a larger zone of the medium shows pink colour, the medium can be restored once by heating for a period not longer than 20 minutes. To avoid excess pressure the bottle should be vented with an aeration cannula during heating as well as cooling down phase

	Safety Data
<b>Toxic ingredients</b>	<ul style="list-style-type: none"> <li>• None</li> </ul>
<b>Basic composition</b>	<ul style="list-style-type: none"> <li>• See typical composition</li> </ul>
<b>Solvent content</b>	<ul style="list-style-type: none"> <li>• None</li> </ul>
<b>Safety data sheet required</b>	<ul style="list-style-type: none"> <li>• Not mandatorily required</li> </ul>